



No. A3-439/23(ii)

Dated: 23.02.2023

### **QUOTATION NOTICE**

Sealed quotations are invited by the undersigned for the supply of Autoclave (Vertical) and Electronic weighing balance as per the specifications and terms & conditions detailed below:

<b>Sl. No.</b>	<b>Name of the equipment</b>	<b>Specifications</b>
1	Autoclave (Vertical)	<ul style="list-style-type: none"><li>• Vertical type</li><li>• Double walled, Inner chamber made of 304 grade SS, Outer SS</li><li>• Lid made of SS should have pressure gauge, safety valve, steam exhaust, vacuum breaker and pedal lifting device</li><li>• Working temperature - 121 °C at 15 or more PSI pressure</li><li>• Chamber volume – 52 litres</li><li>• ISI quality immersion type heater</li><li>• Auto pressure cut off system</li><li>• Digital temperature controller cum indicator</li><li>• Digital timer</li><li>• Inner SS perforated basket</li><li>• Rating : 3 KW</li><li>• Indicators : water level indicators &amp; pressure gauges.</li><li>• Certification – ISO, CE</li></ul>
	Electronic weighing balance	<ul style="list-style-type: none"><li>• Analytical balance with Auto Internal Calibration</li><li>• Including windshield</li><li>• Capacity – 800-1000g</li><li>• Minimum load – 1 mg</li><li>• Readability – 0.001 g</li><li>• Linearity - <math>\pm 0.002</math> g</li><li>• Pan size – 115 mm diameter</li><li>• Response Time – 2-3 Seconds</li><li>• Display – Backlite high resolution graphic display</li><li>• Operating temp - 18 to 30 degree °C</li></ul>

### **Conditions**

1. Quotations are to be addressed to the Professor and Head, Agricultural Research Station, Anakkayam, Anakkayam (PO), Pin- 676509 and should reach the same on or before 11 a.m on **07/03/2023** The quotations received will be opened at 11.30 a.m on the same day in the presence of quotationers or their authorized agents who are present.
2. The quotations should quote the rate at which they are ready to supply the items.
3. The lowest rate quoted, if reasonable, will be accepted and the successful quotationer should supply the materials to the station within 15 days of issue of the supply order.
4. The rate quoted should be inclusive of all taxes, duties, transportation and installation charges, if any.
5. The rate should be valid at least for 6 months from the last date of receipt of quotation.
6. A detailed catalogue regarding the material may be enclosed with quotation.
7. The Trade names, Specification, etc may be specifically indicated.
8. Payment will be effected only after satisfactory installation of equipments with demonstration. The item supplied should be of good quality, otherwise it will be rejected.
9. The undersigned fully reserves the right to accept or reject the quotations without assigning reason thereof
10. All rules and regulations regarding government quotations will be equally applicable to this too .

**Sd/-**

**Dr. Musthafa Kunnathadi**  
Professor & Head

Copy to:

1. Office notice board
2. Circulation
3. KAU Website

